



E.coli Host Cell DNA Residue Detection Kit-QUALITY CONTROL

1. Accuracy

Intra Variation%	4.4-5.6
Inter Variation%	3-15

Standard: $\leq 30\%$

2. Recovery%: 53-81

3. Limit of Quantitation: 30 fg/ μ L