



---

## HEK293 Host Cell DNA Residue Detection Kit-QUALITY CONTROL

### 1. Accuracy

<b>Intra Variation%</b>	10.3-12.3
<b>Inter Variation%</b>	16-19

Standard  $\leq$  30%

2. **Recovery%:** 73-110
3. **Limit of Quantitation:** 30fg/ $\mu$ L