



## Vero Host Cell DNA Residue Detection Kit-QUALITY CONTROL

### 1. Accuracy

<b>Intra Variation%</b>	4.8-12.5
<b>Inter Variation%</b>	14-17

Standard  $\leq$  30%

2. **Recovery%:** 102-117
3. **Limit of Quantitation:** 1fg/ $\mu$ L